

New Delhi, the 30th June, 2006

S.O. 3102.—Whereas the Central Government, after considering the report submitted to it by prescribed authority, is satisfied that the model described in the said report (see the figure given below) is in conformity with the provisions of the Standards of Weights and Measures Act, 1976 (60 of 1976) and the Standards of Weights and Measures (Approval of Models) Rules, 1987 and the said model is likely to maintain its accuracy over periods of sustained use and to render accurate service under varied conditions;

Now, therefore, in exercise of the powers conferred by Sub-sections (7) and (8) of Section 36 of the said Act, the Central Government hereby issues and publishes the certificate of approval of the model of non-automatic weighing instrument (Table top type) with digital indication of special accuracy (Accuracy class I) of series "EHP" and with brand name "EAGLE" (hereinafter referred to as the said model), manufactured by M/s. E.G. Kantawala Pvt. Ltd., Survey No. 28/1, Damodar Nagar, Old Nagar, Mundhawa Road, Kharadi, Pune-411014, Maharashtra and which is assigned the approval mark IND/09/2006/317;



The said model is a electro magnetic force Compensation load cell based non-automatic weighing instrument (Table top type) with a maximum capacity of 4kg. and minimum capacity of 1g. The verification scale interval (e) is 10mg. It has a tare device with a 100 per cent subtractive retained tare effect. The Light Emitting Diode (LED) indicates the weighing results. The instrument operates on 230 volts, 50 Hertz alternative current power supply.

In addition to sealing the stamping plate, sealing shall also be done to prevent the opening of the machine for fraudulent practices and model shall not be changed in terms of its material, accuracy, design, circuit diagram, working principle etc, before or after sale.

Further, in exercise of the powers conferred by Sub-section (12) of Section 36 of the said Act, the Central Government hereby declares that this certificate of approval of the said model shall also cover the weighing instruments of similar make, accuracy and performance of same series with maximum capacity upto 50 kg. with verification scale interval (n) in the range of 50,000 or above for 'e' value of 1mg. or more and with 'e' value of 1×10^k , 2×10^k or 5×10^k , where k is a positive or negative whole number or equal to zero, manufactured by the same manufacturer in accordance with the same principle, design and with the same materials with which, the said approved Model has been manufactured.

[F. No. WM-21(65)/2006]

P. A. KRISHNAMOORTHY, Director of Legal Metrology