

[To be published in the Gazette of India Part II, Section 3, Sub-section (ii)]

Government of India

Ministry of Consumer Affairs, Food and Public Distribution

Department of Consumer Affairs

New Delhi, Dated the 19.9.2016

#### Notification

S.O.-- Whereas the Central Government, after considering the report submitted to it by prescribed authority, is satisfied that the model described in the said report (see the figure given below) is in conformity with the provisions of the Legal Metrology Act, 2009 (1 of 2010) and the Legal Metrology (Approval of Models) Rules, 2011 and the said model is likely to maintain its accuracy over periods of sustained use and to render accurate service under varied conditions;

Now, therefore, in exercise of the powers conferred by section 22 of the Legal Metrology Act, 2009 (1 of 2010) read with sub-rule (6) of rule 8 and sub-rule (4) of rule 11 of the Legal Metrology (Approval of Models) Rules, 2011, the Central Government hereby issues and publishes the certificate of approval of the model of non-automatic weighing instrument with analogue indication of ordinary accuracy (accuracy class-III) of series "ECON" and with brand name "EAGLE" ( hereinafter referred to as the said model), manufactured by M/s E. G. Kantawalla Pvt. Ltd., Survey No 28/1, Damodar Nagar, Old Nagar, Mundhwa Road, Kharadi, Pune-411014 and which is assigned the approval mark IND/09/16/ 414 ;

The said model is a spring and lever based mechanical non-automatic weighing instrument with a maximum capacity of 130kg and minimum capacity of 10kg. The verification scale interval (e) is 1kg. The pointer on the graduated dial indicates the results of the measurements. It is used as mechanical person weighing machine.

Figure- 1 Model






Figure- 2 Schematic diagram of the sealing arrangement

The sealing provision for the weighing instrument is provided by two body fitting screws as shown above in figure 2,

Further, in exercise of the power conferred by sub-rule (9) of rule 8 of the Legal Metrology (Approval of Models) Rules, 2011, the Central Government hereby declares that this certificate of approval of the said model shall also cover the weighing instruments of similar make, accuracy and performance of same series with maximum capacity up to 200kg with number of verification scale interval (n) in the range of 100 to 1,000 for 'e' value of 5g or more and with 'e' value of  $1 \times 10^k$ ,  $2 \times 10^k$  or  $5 \times 10^k$ , where k is a positive or negative whole number or equal to zero manufactured by the same manufacturer in accordance with the same principle, design and with the same materials with which, the said approved model has been manufactured.

[F.No.WM-21(275)/2016]

*Bahar Nishad*

(B.N Dixit)

Director (Legal Metrology) to Government of India